

# EU Type Examination Certificate

This is to certify that:

GVS Filter Technology UK Limited  
NFC House  
Vickers Industrial Estate  
Mellishaw Lane  
Morecambe  
Lancashire  
LA3 3EN  
United Kingdom

Holds Certificate Number:

CE 711763

In respect of:

**Filtering masks to EN 149:2001+A1:2009**  
**See Continuation sheets for details.**


on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified  
Body for the above Regulation  
(Notified Body Number 2797):

Previous Notified Body: BSI 0086

First Issued: 2019-05-01

Latest Issue: 2020-06-04



Drs. Dave Hagenaaars, Managing Director

Effective Date: 2020-06-04

Expiry Date: 2024-05-01

Page: 1 of 4



...making excellence a habit.™

# EU Type Examination Certificate

No. CE 711763

## Product Specification

### Vertical Fold Flat Range with clogging

**Description:** Vertical fold flat disposable face mask with clogging option 7.17d

**Technical specification:** **EN 149:2001+A1:2009** Respiratory protective devices – Filtering half masks to protect against particles.  
Classification – Reusable – R or Single shift - NR

**FFP Class:** **FFP1 R D**

**Model:**  
CN P1 F10000  
CN P1V F10050 (mask with exhalation valve fitted)

**FFP Class:** **FFP1 NR D**

**Model:**  
CN P1C F10005 (mask with activated charcoal)  
CNP1CV F10055 (mask with activated charcoal and exhalation valve fitted)

**FFP Class:** **FFP2 R D**

**Model:**  
CN P2 F20000  
CN P2V F20050 (mask with exhalation valve fitted)

**FFP Class:** **FFP2 NR D**

**Model:**  
CN P2CV F20055 (mask with activated charcoal and exhalation valve fitted)

First Issued: 2019-05-01

Latest Issue: 2020-06-04

Effective Date: 2020-06-04

Expiry Date: 2024-05-01

Page: 2 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.  
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands  
A member of BSI Group of Companies.

# EU Type Examination Certificate

No. CE 711763

## Product Specification

**Description:** Vertical fold flat disposable face mask with clogging option 7.17d

**Technical specification:** **EN 149:2001+A1:2009** Respiratory protective devices – Filtering half masks to protect against particles.  
Classification – Reusable – R or Single shift - NR

**FFP Class:** **FFP3 R D**

**Model:** CN P3 F30000  
CN P3V F30050 (mask with exhalation valve fitted)  
CN P3 F31000

**FFP Class:** **FFP3 NR D**

**Model:** CN P3CV F30055 (mask with activated charcoal and exhalation valve fitted)

First Issued: 2019-05-01  
Latest Issue: 2020-06-04

Effective Date: 2020-06-04  
Expiry Date: 2024-05-01

# EU Type Examination Certificate

No. CE 711763

## Certificate Administration Details

Technical File Reference: Particle Filtering Half Masks EN149 Cap.02/Rev. 0

## Certificate Amendment Record

Issue date	Comments	BSI Internal Project No.
June 2020	Addition of model F31000 and Module C2 reference for GVS Italy, Bologna, Sud and GVS China, Suzhou and GVS Romania, Ciorani. Italcert C2 up-issued to PPE194AC379.	2797:20:3174968
April 2020	Addition of Italcert Module C2 reference for GVS Italy, Bologna.	2797:20:3166806
May 2019	First issue. Transition from certificate CE 672393 to the PPE Directive 89/686.	0086:19:9690679

**Note:** The Certificate holder is responsible for keeping the Notified Body advised of changes to any aspect of the overall process used in the manufacture of the product.

## Monitoring of manufactured PPE.

The EC Quality of Production by Monitoring requirements, Module C2 and D, for the product are referenced in the issued Certificates:

GVS Site	Notified Body	Module	Certificate	Products
NFC House, UK	BSI	D	CE 672394	All listed on CE 711763
Italy, Bologna	Italcert	C2	PPE194AC379	CN P3 F30000
Italy, Bologna	BSI	C2	CE 727721	CN P3 F31000
Italy, Sud	Italcert	C2	PPE194AC379	CN P3 F30000
Italy, Sud	BSI	C2	CE 727721	CN P3 F31000
China, Suzhou	Italcert	C2	PPE194AC379	CN P3 F20000
Romania, Ciorani	Italcert	C2	PPE194AC379	CN P3 F30000

First Issued: 2019-05-01

Latest Issue: 2020-06-04

Effective Date: 2020-06-04

Expiry Date: 2024-05-01