

UKCA Type Examination Certificate

This is to certify that:

3M United Kingdom Plc
3M Centre
Cain Road
Bracknell
RG12 8HT
United Kingdom

Holds Certificate Number:

UKCA 748220

In respect of:

Respiratory protective devices to EN 149:2001+A1:2009

Filtering half masks to protect against particles

Models: 3M™ Aura™ Particulate Respirators

9310+, 9312+, 9320+, 9322+, 9330+, 9332+, 1872V+ and 1873V+.

On the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation 2016/425 on Personal Protective Equipment, as amended to apply in GB (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II.

For and on behalf of BSI, an Approved Body for the above Regulation
(Approved Body Number 0086):



Shahm Barhom, Group Product Certification Director

First Issued: 2021-08-16

Latest Issue: 2023-01-06

Effective Date: 2023-01-06

Expiry Date: 2026-08-15



Page: 1 of 3

UKCA Type Examination Certificate

No. UKCA 748220

Product Specification

| | | |
|---------------|---|--------------------------|
| Product Name: | 3M™ Aura™ Particulate Respirators | |
| Product Type: | Filtering half masks to protect against particles | |
| Models: | Particulate Respirators: | Health Care Respirators: |
| | 9310+ | 1872V+ |
| | 9312+ | 1873V+ |
| | 9320+ | |
| | 9322+ | |
| | 9330+ | |
| | 9332+ | |

Technical Specification: Designated Standard: EN 149:2001+A1:2009

Product Description: The particulate respirators are designed to protect against solid and non-volatile liquid particles. The masks are held on the face by a pair of elasticated straps. Some models incorporate a single exhalation valve. The models are single shift devices (denoted by the classification symbol NR) and have the Dolomite Clogging option (denoted by the classification symbol D).

| EN 149 Classification: | Model | Valved | EN149 Classification |
|------------------------|--------|------------|----------------------|
| | 9310+ | Not valved | FFP1 NR D |
| | 9312+ | With valve | FFP1 NR D |
| | 9320+ | Not valved | FFP2 NR D |
| | 9322+ | With valve | FFP2 NR D |
| | 9330+ | Not valved | FFP3 NR D |
| | 9332+ | With valve | FFP3 NR D |
| | 1872V+ | With valve | FFP2 NR D |
| | 1873V+ | With valve | FFP3 NR D |

First Issued: 2021-08-16
Latest Issue: 2023-01-06

Effective Date: 2023-01-06
Expiry Date: 2026-08-15

UKCA Type Examination Certificate

No. UKCA 748220

Packaged variants

In addition to the products referenced on this Certificate the products may also be sold as market specific packaged variants. Such variants will be differentiated and the Technical File(s) will be updated with the appropriate information.

Kits and packouts

The products referenced on this Certificate may also be combined with other 3M products into a kit or packout. There will be no change to the product but the User Information may vary, in such instances the applicable Technical File(s) will be updated with the appropriate information.

Certificate Administration Details

Technical File Reference: TF0478

Certificate Amendment Record:

| Issue date | Comments | BSI Review No. |
|--------------|--|-----------------|
| August 2021 | First issue | 0086:21:3422821 |
| January 2023 | The removal of the 3M™ Aura™ Health Care Respirators – Fluid Resistant models: 1861+, 1862+, 1863+, 1883+. | 0086:23:3793636 |

Certificate validity

The Certificate holder is responsible for ensuring that the Approved Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate for the products is also dependent on the maintenance of the UKCA Conformity to Type based on Quality Assurance of the Production Process, Annex VIII (Module D), as referenced on BSI issued Certificate UKCA 745215.

First Issued: 2021-08-16

Latest Issue: 2023-01-06

Effective Date: 2023-01-06

Expiry Date: 2026-08-15

Page: 3 of 3

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.